

LECHPOL ELECTRONIC S.R.L.
Resita, B-dul Republicii 5, ap.4
Caras-Severin
J11/1142/2007
RO22976532
Tel. 0359/453700

**DECLARATIE
DE
CONFORMITATE**

Noi, S.C. LECHPOL ELECTRONIC S.R.L., cu sediul in Resita, judetul Caras-Severin, B-dul Republicii 5, ap. 4, cu Certificat de inregistrare nr. J11/1142/2007, RO22976532, declaram că produsul:

MULTIMETRU REZISTENTA IZOLATIE UT502A UNI-T (cod MIE0294)

la care se refera aceasta declaratie, respecta cerintele esentiale si alte prevederi pertinente ale directivei 2004/108/EC.

Declaratia de conformitate a fost intocmita pe baza documentatiei obtinuta de la furnizor.

LECHPOL ELECTRONIC SRL
Dan Pirtea

Sintandrei,
8 sep. 2008



Test Verification of Conformity

On the basis of the referenced test report(s), the sample(s) of the below product has been found to comply with the relevant harmonized standard(s) to the directive(s) listed on this verification at the time the tests were carried out. The manufacturer may indicate compliance to said directive(s) by signing a DoC himself and applying the CE-marking to products identical to the tested sample(s). In addition, the manufacturer shall file and keep the documentation according to the rules of the applicable directive(s) and shall consider changes of the standard(s) if relevant. Additional requirements may be applicable such as additional directives or local laws.


Applicant Name & Address	: Uni-Trend Group Limited Rm 901, 9/F, Nanyang Plaza, 57 Hung To Road, Kwun Tong, Hong Kong
Product(s) Tested	: Insulation resistance tester
Ratings and principal characteristics	: DC 9V battery. Measurement Category: CAT III 500 V, CAT II 1000 V
Model(s)	: UT501, UT502
Brand name	: UNI-T
Relevant Standard(s) / Specification(s) / Directive(s)	: EN 61326-1: 2006/ Electrical equipment for measurement, control and laboratory use – EMC requirements Part 1: General requirements EN 61326-2-2: 2006/ Electrical equipment for measurement, control and laboratory use – EMC requirements Part 2-2: Particular requirements – Test configuration, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems EMC Directive 2004/108/EC
Verification Issuing Office Name & Address	: Same as Legal Entity
Verification/Report Number(s)	: GZ08070301-1/GZ08070301-1

Note 1: This verification is part of the full test report(s) and should be read in conjunction with it.

This Verification is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this Verification. Only the Client is authorized to copy or distribute this Verification. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results referenced from this Verification are relevant only to the sample tested. This Verification by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.



Signature


Name: Yannie Wang
Position: Assistant Manager
Date: 08 September, 2008